

2023 年 12 月 8 日科研楼实验室安全隐患排查、卫生
巡查异常情况通报

2023 年 12 月 8 日科研中心教职工对第五实验楼、第二实验楼五
六楼、第四实验楼一二楼科研实验室进行安全隐患排查及卫生检查，
检查结果如下：

序 号	日期	存在问 题	所在位置	照片
1	12.8	垃圾满 未处理	第四实验楼 212	
2	12.8	干燥箱 使用未 登记	第五实验楼 106	

3	12.8	实验室 有水杯	第五实验楼 612	 A photograph showing a green thermos cup on a dark desk in a laboratory setting. The cup is upright and has a white lid. A pair of yellow-handled pliers is lying on the desk next to it. In the background, there is a computer monitor and some papers. A red text label at the bottom of the image reads "第五实验楼612".
4	12.8	钢瓶未 固定	第五实验楼 515	 A photograph of a tall, silver gas cylinder standing upright in a laboratory. The cylinder is not secured to a wall or stand. It is positioned next to a white wall and a blue carpeted floor. A red text label at the bottom of the image reads "第五实验楼515".
5	12.8	实验室 有水杯	第五实验楼 413	 A photograph of a green thermos cup on a green desk in a laboratory. The cup is upright and has a black lid. In the background, there is a window and a red box. A red text label at the bottom of the image reads "第五实验楼413".